



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE
United States Patent and Trademark Office
Address: COMMISSIONER FOR PATENTS
P.O. Box 1450
Alexandria, Virginia 22313-1450
www.uspto.gov

5W

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/017,455	12/13/2001	Densen Cao	5061.3 P	6763

7590 04/13/2004

Daniel McCarthy
PARSONS, BEHLE & LATIMER
201 South Main Street, Suite 1800
P.O. Box 45898
Salt Lake City, UT 84145-0898

EXAMINER

LEWIS, RALPH A

ART UNIT	PAPER NUMBER
----------	--------------

3732

DATE MAILED: 04/13/2004

12

Please find below and/or attached an Office communication concerning this application or proceeding.

Office Action Summary

Application No.

10/017,455

Applicant(s)

CAO, DENSEN

Examiner

Ralph A. Lewis

Art Unit

3732

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If the period for reply specified above is less than thirty (30) days, a reply within the statutory minimum of thirty (30) days will be considered timely.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

Status

- 1) ☒ Responsive to communication(s) filed on 29 December 2003.
- 2a) ☒ This action is **FINAL**. 2b) ☐ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

Disposition of Claims

- 4) ☒ Claim(s) 1-5, 7-11, 13-22 and 24 is/are pending in the application.
- 4a) Of the above claim(s) _____ is/are withdrawn from consideration.
- 5) ☒ Claim(s) 1-5, 7-11, 13-17 and 24 is/are allowed.
- 6) ☒ Claim(s) 18-22 is/are rejected.
- 7) ☐ Claim(s) _____ is/are objected to.
- 8) ☐ Claim(s) _____ are subject to restriction and/or election requirement.

Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☐ The drawing(s) filed on _____ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All b) ☐ Some * c) ☐ None of:
- ☐ Certified copies of the priority documents have been received.
 - ☐ Certified copies of the priority documents have been received in Application No. _____.
 - ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

* See the attached detailed Office action for a list of the certified copies not received.

Attachment(s)

- ☒ Notice of References Cited (PTO-892)
- ☐ Notice of Draftsperson's Patent Drawing Review (PTO-948)
- ☐ Information Disclosure Statement(s) (PTO-1449 or PTO/SB/08)
Paper No(s)/Mail Date _____
- ☐ Interview Summary (PTO-413)
Paper No(s)/Mail Date. _____
- ☐ Notice of Informal Patent Application (PTO-152)
- ☐ Other: _____

Rejections based on Prior Art

The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless –

(a) the invention was known or used by others in this country, or patented or described in a printed publication in this or a foreign country, before the invention thereof by the applicant for a patent.

(b) the invention was patented or described in a printed publication in this or a foreign country or in public use or on sale in this country, more than one year prior to the date of application for patent in the United States.

Claims 18 and 21 are rejected under 35 U.S.C. 102(a) as being anticipated by Mills (WO 99/16136).

Mills discloses a dental curing light (page 1, second paragraph) comprised of a hand held wand (Figure 5) having batteries 54, a light module 47, an elongated heat sink 45, 50, 51, having a distal end surface serving as a mounting platform on which primary heat sink 48 is mounted and light emitting semiconductors 43 mounted to the primary heat sink 48. In regard to claims 18 and 21, it is presumed that the Mills device inherently must contain an on/off switch so that the device may be operated in a normal manner, even though Mills fails to explicitly identify such a conventional component.

The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

Claims 18 and 21 are rejected under 35 U.S.C. 103(a) as being unpatentable over Mills (WO 99/16136)

In regard to claims 3 and 9, operation of the Mills device within conventional current levels would have been obvious to one of ordinary skill in the art. In regard to claims 4 and 10, Mills discloses that a fan 49 may be used to help cool the device (note page 14), merely providing for vents through which the fanned air may circulate to cool the device would have been obvious to one of ordinary skill in the art. In regard to claims 18 and 21, Mills does not explicitly appear to state that the disclosed dental photo curing device has an on/off switch (i.e. "controls for initiating and terminating light transmission" and "circuitry in electrical connection with said controls" as required by claim 18). The use, however, of a conventional on/off switch to turn the device on and off when being used would have most certainly been obvious to the ordinarily skilled artisan.

Claims 19, 20 and 22 are rejected under 35 U.S.C. 103(a) as being unpatentable over Mills (WO 99/16136) in view of Doiron et al (5,698,866).

In Mills the LEDs are mounted directly on a flat heat sink 48. Doiron et al, however, teach that an improvement over mounting diodes on a flat surface (Figures 9 and 10) is mounting them in a well (Figures 11 and 12) formed on the heat sink so that more light from the LEDs is reflected forward in the desired direction. To have mounted

the Mills LEDs in wells as taught by Doiron et al so that more light is reflected forward in the desired direction would have been obvious to one of ordinary skill in the art.

In regard to claim 15, Mills does not expressly state how the LEDs 43 are connected to platform 48, however, the use of a conventional prior art adhesives would have been obvious to one of ordinary skill in the art, it is further noted that all adhesives have at least some degree of heat conductivity and light reflectivity.

Allowable Subject Matter

Claims 1-5, 7-11, 13-17 and 24 are allowed.

Action Made Final

THIS ACTION IS MADE FINAL. Applicant is reminded of the extension of time policy as set forth in 37 CFR 1.136(a).

A shortened statutory period for reply to this final action is set to expire THREE MONTHS from the mailing date of this action. In the event a first reply is filed within TWO MONTHS of the mailing date of this final action and the advisory action is not mailed until after the end of the THREE-MONTH shortened statutory period, then the shortened statutory period will expire on the date the advisory action is mailed, and any extension fee pursuant to 37 CFR 1.136(a) will be calculated from the mailing date of the advisory action. In no event, however, will the statutory period for reply expire later than SIX MONTHS from the mailing date of this final action.


Art Unit: 3732

Prior Art

Melikechi et al (6,511,317), Otsuka (6,638,063), Decaudin et al (6,692,250), Logan et al (6,692,251), Plank (6,695,614), and Fischer et al (6,702,576) are made of record.

Any inquiry concerning this communication should be directed to **Ralph Lewis** at telephone number **(703) 308-0770**. Fax (703) 872-9306. The examiner works a compressed work schedule and is unavailable every other Friday. The examiner's supervisor, Kevin Shaver, can be reached at (703) 308-2582.

R.Lewis
April 5, 2004



Ralph A. Lewis
Primary Examiner